



December 21, 2018

Tissue Regeneration Technologies, LLC
% Jennifer Daudelin, MSJ
Regulatory Consultant III
M Squared Associates, Inc.
575 Eighth Avenue, Suite 1212
New York, New York 10018

Re: K182682

Trade/Device Name: OrthoGold 100™
Regulation Number: 21 CFR 890.5660
Regulation Name: Therapeutic Massager
Regulatory Class: Class I
Product Code: ISA
Dated: September 26, 2018
Received: September 26, 2018

Dear Jennifer Daudelin:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/CombinationProducts/GuidanceRegulatoryInformation/ucm597488.htm>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/>) and CDRH Learn (<http://www.fda.gov/Training/CDRHLearn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<http://www.fda.gov/DICE>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,


Vivek J. Pinto -S

for Carlos L. Peña, PhD, MS
Director
Division of Neurological
and Physical Medicine Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K182682

Device Name

OrthoGold 100

Indications for Use (Describe)

The OrthoGold 100 is intended for the activation of connective tissue.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary
Tissue Regeneration Technologies, LLC
OrthoGold 100™

The following information is provided as required by 21 CFR § 807.87 for the Tissue Regeneration Technologies, LLC 510(k) premarket notification. In response to the Safe Medical Devices Act of 1990, the following is a summary of the safety and effectiveness information upon which the substantial equivalence determination is based.

Sponsor: Tissue Regeneration Technologies, LLC
251 Heritage Walk
Woodstock, GA 30188

Contact: Jennifer A. Daudelin, M.S.J.
M Squared Associates, Inc.
575 Eighth Avenue, Suite 1212
New York, NY 10018
Ph: 703-562-9800 x251
Fax: 703-562-9797
Email: jdaudelin@msquaredassociates.com

Date Prepared: September 24, 2018

Proposed Class: I

Proprietary Name: OrthoGold 100™

Common Name: Therapeutic Massager

Classification Name: Massager, Therapeutic, Electric

Regulation Number: 21 CFR 890.5660

Product Codes: ISA

Predicate Device(s):

Manufacturer	Device Name	510(k) Number	Procode	Class
Asclepion Laser Technologies GmbH	Dermablade Effect	K081541	ISA	I

Indications for Use

The OrthoGold 100™ is intended for the activation of connective tissue.

Device Description

The OrthoGold 100™ is a pulsed acoustic wave device. It includes an electrically powered generator to generate transient compressed air that rapidly expands to create the acoustic waves, which in turn are propagated through a water-filled coupling membrane attached to the hand-held applicator. The hand-held applicator reflects the acoustic waves towards the treatment area through a silicone membrane and ultrasound transmission gel.

Performance Data

Verification and validation testing was performed and demonstrated that the OrthoGold 100™ meets the design specifications and is safe and effective for its intended use. All tests required by the verification and validation plan were completed and passed. The OrthoGold 100™ software was validated and demonstrated to be of a Moderate level of concern; while hazard analysis / risk management was performed and demonstrated that all risks are mitigated to an acceptable level. The OrthoGold 100 was tested and demonstrated to conform to the general safety requirements of IEC 60601-1:2005; as well as the electromagnetic compatibility requirements of IEC 60601-1-2:2014 (4th Ed.) and 60601-2-36. In-vitro testing was performed to determine applicator displacement, force and penetration depth and was demonstrated to be equivalent to the AW module of the Dermablate predicate device. In addition, probe cover testing and transport verification and validation was also conducted. The performance testing demonstrated that the OrthoGold 100 is substantially equivalent to the predicate device.

The table below compares the OrthoGold 100 characteristics to the predicate device.

Product Characteristic	Subject Device OrthoGold 100™	Predicate Device Dermablate Effect	Comparison
510(k) Number	To be assigned	K081541	NA
Indications for Use	Activation of connective tissue	Activation of connective tissue	Identical
Modes of Action	Unfocused pressure pulses	Radial (unfocused) pressure waves, or extracorporeal pulse activation respectively	Similar
Mechanisms of Action	Extracorporeally induced unfocused pressure pulses	Pneumatically generated vibrations + unfocused pressure pulses	Equivalent
Maximum and Minimum intensity settings	1 to 16	Specific Value Not Available	Similar

Product Characteristic	Subject Device OrthoGold 100™	Predicate Device Dermablate Effect	Comparison
Number and size of treatment applicator heads	OP155 Size: 230 x ø 70 mm	Specific Value Not Available	Similar
Maximum and minimum displacements of applicator heads	Not Applicable	Specific Value Not Available	NA
Type of application (e.g., continuous vibration at a fixed frequency);	Continuous at various frequencies	Specific Value Not Available	Similar
Maximum and minimum vibration frequency	Frequency of 1 - 8 Hz in steps of 0.5 Hz	Specific Value Not Available	Similar
Driving Power	High voltage 2 - 7 kV Capacitor: 0,2 uF	Specific Value Not Available	Similar
Power Supply	115 VAC	Specific Value Not Available	Similar
Maximum applicator force	Not Applicable	Specific Value Not Available	NA
Maximum applicator displacement	Not Applicable	Specific Value Not Available	NA
Maximum penetration depth	25.4 mm at energy level 16	30mm (3cm)	Similar
Energy flow density PIIT [mJ/mm2]	0.00017 - 0.04403 at energy level 1 - 16	0.018-0.25	Similar
Operating mode	Continuous	Specific Value Not Available	Similar
Projectile mass (g)	Not Applicable	Specific Value Not Available	NA
Pulse repeat rate (1/s)	1 - 8 Hz	1-11 Hz	Similar
Number of pulses (min and max)	500 - 2000	Specific Value Not Available	Similar
Maximum operating temperature	Room temperature	Specific Value Not Available	Similar
Type of acoustic wave generation	Electro hydraulic, spark gap under water caused by discharge of high voltage condensers	Ballistic technology, pressurized air pulses accelerate a projectile within a guiding tube	Similar
Peak compressional acoustic pressure pc [Mpa]	9.27 at energy level 16	13.4	Similar
Peak rarefactional acoustic pressure pcr[Mpa]	-1.52 at energy level 16	Specific Value Not Available	Similar
Description of the spatial distribution of the acoustic pressure and intensity	Unfocused acoustic pressure field, see pressure measurements	Specific Value Not Available	Similar
Positive peak pressure amplitude (MPa) pc [Mpa]	0.43 - 9.27 at energy level 1 - 16	13.4	Similar

Product Characteristic	Subject Device OrthoGold 100™	Predicate Device Dermablate Effect	Comparison
Negative peak pressure amplitude (MPa) p _{cr} [Mpa]	-0.17 to -1.52 at energy level 1 - 16	Specific Value Not Available	Similar
Derived focal acoustic pulse energy (mJ) EbT [mJ]	0.022 - 2.278 at energy level 1 - 16	Specific Value Not Available	Similar
Derived pulse intensity integral, integrated over total temporal integration limits PIIT [mJ/mm ²]	0.00017 - 0.04403 at energy level 1 - 16	Specific Value Not Available	Similar
Rise time (ns) (10% - 90%) t _r [us]	1.89 - 0.28 at energy level 1 - 16	Specific Value Not Available	Similar
Compressional pulse duration (μs) t _{FWHMpc} [uS]	1.23 - 0.77 at energy level 1 - 16	Specific Value Not Available	Similar

Technological Characteristics and Substantial Equivalence

The OrthoGold 100™ has the same indications for use and similar design features as compared with the predicate system. The bench testing demonstrates that the performance characteristics of the OrthoGold 100™ are equivalent to those of other legally marketed therapeutic massagers, and therefore supports a determination of Substantial Equivalence for the proposed indications for use. Any differences between the subject and predicate device would not render the device NSE, affect the safety or effectiveness, or raise different questions of safety and effectiveness.